



TO: Nebraska Healthcare Providers, Laboratories, Health Care Facilities, Infection Control,

Hematology, ERs, Critical Care, Oncology, Pathologists, and Public Health

FROM: Thomas J. Safranek, M.D. Thomas Williams, M.D.

State Epidemiologist Chief Medical Officer

402-471-2937 PHONE Director, Division of Public Health,

402-471-3601 FAX Department of Health and Human Services

RE: Multi-state outbreak of Serratia marcescens bloodstream infections

DATE: April 30, 2018

Background:

CDC, state and local health departments, and FDA are investigating an ongoing multi-state outbreak of *Serratia marcescens* bloodstream infections. Molecular testing of clinical isolates from patients in 3 states (CO, TN, and OH) has resulted in an indistinguishable or closely related pulsed-field gel electrophoresis (PFGE) pattern, suggesting a possible shared exposure. The infections are primarily among pediatric patients with central venous access. Common exposures identified include heparin and saline flushes, specifically, PosiFlush Pre-Filled Heparin Lock Flush Syringes (10 USP units/mL and 100 USP units/ml) and PosiFlush Sterile Field Saline Flush Syringes, all manufactured by BD. No products have tested positive for *S. marcescens* to date. Product testing is ongoing; other products cannot be ruled out at this time. Out of an abundance of caution, BD is voluntarily recalling certain lots of BD PosiFlushTM Heparin Lock Flush and BDTM Pre-Filled Normal Saline Flush syringes due to this potential contamination with *S. marcescens*.

Additional information and a list of the recalled products and lot numbers can be found here: http://www.bd.com/en-us/support/alerts-and-notices/medical-device-recall-for-bd-posiflush-heparin-lock-flush-syringes-and-bd-pre-filled-normal-saline-syringes.

Recommendations:

- Clinicians, laboratories, and healthcare facilities should notify Dr. Maureen Tierney or Dr. Caitlin Pedati when *S. marcescens* is detected from a sterile site specimen (e.g., blood).
- Nebraska laboratories are requested to hold *S. marcescens* isolates (especially from children) from sterile site specimens for potential follow-up testing at the NPHL.
- Clinicians and healthcare facilities that use PosiFlush Pre-Filled Heparin Lock Flush Syringes (10 USP units/mL and 100 USP units/ml) or PosiFlush Sterile Field Saline Flush Syringes manufactured by BD might consider using alternative products, if possible.

This investigation is ongoing. If you have any questions or concerns, please contact:

Dr. Maureen Tierney Dr. Caitlin Pedati

Healthcare-Associated Infections Director, NDHHS Medical Epidemiologist, NDHHS

Office: 402-471-6549 Office: 402-471-9148

Email: Maureen. Tierney@nebraska.gov Email: caitlin.pedati@nebraska.gov